



EMI/EMS Test Result
According to IEC60601-1-2 4th Edition (EMS)

Model Name : AME600F series

Approved : Satoshi Uetani

The EUT is operated with following condition during EMI/EMS test.

Input Voltage : 230, 240VAC / 50Hz
Output Current : Rated Current
Ambient Temperature : 25°C ± 10°C

Prepared : Enkyo Kaku

#	Subject	Reference standard	Test Condition	Criteria *1	Result	
1	EMI		Conducted Emission	EN55011, EN55032 Class B CISPR 32 Class B FCC Part15 Class B VCCI Class B	-	Pass
			Radiated Emission	EN55011, EN55032 Class B CISPR 32 Class B FCC Part15 Class B VCCI Class B	-	Pass
			Harmonic Current	IEC61000-3-2	Class A	-
4	Electrostatic discharge immunity test	IEC61000-4-2	Contact Discharge : Level 4 (8kV) Air Discharge : Level 4 (15kV) Applied to Input, Output, FG and Chassis	A	Pass	
5	Radiated, radio-frequency, electromagnetic field immunity test	IEC61000-4-3	10V/m : (80MHz~2.7GHz) 80% Amplitude modulated	A	Pass	
6	Electrical fast transient / Burst immunity test	IEC61000-4-4	Level 4 (4kV) Repetition Rate : 5kHz and 100kHz	A	Pass	
7	Surge immunity test	IEC61000-4-5	Line to Line : Level 3 (1kV) Line to Earth : Level 4 (4kV)	A	Pass	
8	Immunity to conducted disturbances, induced by radio-frequency fields	IEC61000-4-6	Voltage Level (e.m.f.) : Level 3 (10Vrms)	A	Pass	
9	Power frequency magnetic field Immunity test	IEC61000-4-8	Magnetic Field Strength : Level 4 (30A/m)	A	Pass	
10	Voltage dips, short interruptions and voltage variations immunity test	IEC61000-4-11	(1) 100% dip for 10ms, 0°, 45°, 90°, 135°, 180°, 225°, 270° and 315°	A	Pass	
			(2) 100% dip for 20ms, 0°	A *2	Pass	
			(3) 60% dip for 100ms, 0°	A *3	Pass	
			(4) 30% dip for 500ms, 0°	A	Pass	
			(5) 100% dip for 5 seconds (short interruption)	B	Pass	

*1 Definition of Criteria

Criteria A : (1) No output voltage drop with control circuit failure.
(2) No protection circuit and other circuit malfunction.

Criteria B : (1) Temporary loss of functionality or degradation of performance result during or after testing, but EUT recovers its normal performance without operator intervention.
(2) No protection circuit and other circuit failure.

*2 Output Current : 80% or less of rated current (at 100VAC)
Output Current : 80% or less of rated current (at 240VAC)

*3 Output Current : 70% or less of rated current (at 100VAC)

<Notes>

Power supply shall not determine the final equipment performance against EMS test. Therefore we confirmed the output voltage performance only. EMS test should be performed as a final product.