



Ref. Certif. No.

**US-42719-UL**

**IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME**

**CB TEST CERTIFICATE**

Product	Switching Power Supply for use in medical equipment
Name and address of the applicant	Cosel Co., Ltd. 1-6-43 Kamiakae-machi Toyama-shi, Toyama 930-0816 Japan
Name and address of the manufacturer	Cosel Co., Ltd. 1-6-43 Kamiakae-machi Toyama-shi, Toyama 930-0816 Japan
Name and address of the factory	Cosel Co., Ltd. Tateyama Factory 78 Dogenji, Tateyamamachi Nakaniikawa-gun, Toyama 930-0241 Japan <input type="checkbox"/> Additional Information on page 2
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	Input: AC 100-240 V, 50-60 Hz, 3.3 A Output: See CB Test Report for details.
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	CTF Stage 2
Model / Type Ref.	GMA300F-12, GMA300F-15, GMA300F-24, GMA300F-30, GMA300F-48, GMA300F-56 May be followed by -#####. Where # is J1, J3, R3, C, I2 or blank
Additional information (if necessary may also be reported on page 2)	<b>Additionally evaluated to:</b> EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021  Class I  The risk management requirements of the standard were not addressed  National Differences: EU Group Differences, CA, US <input type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	E161890-D1052-1/A0/C0-CB issued on 2023-08-29

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2023-09-11

Signature:

Jofanta M. Wroblewska



Ref. Certif. No.

US-42697-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product	Switching Power Supply for use in medical equipment
Name and address of the applicant	Cosel Co., Ltd. 1-6-43 Kamiakae-machi Toyama-shi, Toyama 930-0816 Japan
Name and address of the manufacturer	Cosel Co., Ltd. 1-6-43 Kamiakae-machi Toyama-shi, Toyama 930-0816 Japan
Name and address of the factory	Cosel Co., Ltd. Tateyama Factory 78 Dogenji, Tateyamamachi Nakaniikawa-gun, Toyama 930-0241 Japan <input type="checkbox"/> Additional Information on page 2
Note: When more than one factory, please report on page 2	
Ratings and principal characteristics	Input: AC 100-240 V, 50-60 Hz, 3.3 A Output: See CB Test Report for details.
Trademark / Brand (if any)	
Customer's Testing Facility (CTF) Stage used	CTF Stage 2
Model / Type Ref.	GMA300F-12, GMA300F-15, GMA300F-24, GMA300F-30, GMA300F-48, GMA300F-56 May be followed by -#####. Where # is J1, J3, R3, C, E, I2 or blank
Additional information (if necessary may also be reported on page 2)	<b>Additionally evaluated to:</b> EN 60601-1:2006, EN 60601-1:2006/A1:2013, EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021  Class II  The risk management requirements of the standard were not addressed  National Differences: EU Group Differences, CA, US <input type="checkbox"/> Additional Information on page 2
A sample of the product was tested and found to be in conformity with	IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012, IEC 60601-1:2005/AMD2:2020
As shown in the Test Report Ref. No. which forms part of this Certificate	E161890-D1052-1/A0/C0-CB issued on 2023-08-29

This CB Test Certificate is issued by the National Certification Body



- UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL Solutions (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see [www.ul.com/ncbnames](http://www.ul.com/ncbnames)

Date: 2023-09-11

Signature:

Jofanta M. Wroblewska