



Ref. Certif. No.

DK-148097-UL

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Switching Power Supply

Name and address of the applicant

WUXI COSEL ELECTRONICS CO LTD
NO.82-9 · DICUI ROAD
BINHU DISTRICT Wuxi, Jiangsu, 214072
China

Name and address of the manufacturer

WUXI COSEL ELECTRONICS CO LTD
NO.82-9 · DICUI ROAD
BINHU DISTRICT Wuxi, Jiangsu, 214072
China

Name and address of the factory

WUXI COSEL ELECTRONICS CO LTD
NO.82-9 · DICUI ROAD
BINHU DISTRICT Wuxi, Jiangsu, 214072
China

Note: When more than one factory, please report on page 2

☐ Additional Information on page 2

Ratings and principal characteristics

Model UMA60F-5,
Input rated 100-240VAC, 0.7A, 50-60Hz, Output rated 5 VDC/Max 6.0A
☒ Additional Information on page 2

Trademark / Brand (if any)



Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

UMA60F-5, UMA60F-7R5, UMA60F-12, UMA60F-15, UMA60F-24
☒ Additional Information on page 2

Additional information (if necessary may also be reported on page 2)

Additionally evaluated to: EN 60601-1:2006, EN 60601-1:2006/A1:2013,
EN 60601-1:2006/A12:2014, EN 60601-1:2006/A2:2021
National Differences: EU Group Differences, CA, US
The risk management requirements of the standard were not addressed
☐ Additional Information on page 2

A sample of the product was tested and found to be in conformity with

IEC 60601-1:2005, IEC 60601-1:2005/AMD1:2012,
IEC 60601-1:2005/AMD2:2020

As shown in the Test Report Ref. No. which forms part of this Certificate

E510569-D1005-2/A0/C0-CB issued on 2023-12-11

This CB Test Certificate is issued by the National Certification Body



☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
☒ UL Solutions (Denko), Borupvang 5A DK-2750 Ballerup, DENMARK
☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2023-12-20

Signature:

Thomas Wilson



Ref. Certif. No.

DK-148097-UL

Additional Model Detail(s):

UMA60F-5, UMA60F-7R5, UMA60F-12, UMA60F-15, UMA60F-24, UMA60F-36, UMA60F-48, may be provided with suffix "-wxyz", where "w" = "E" or blank, "x" = "J4" or "T" or "J3" or blank, "y" = "C" or blank, "z" = "Y" or blank. the last one "-" can be excluded when the variable "w", "x", "y", "z" in blank

Additional Ratings:

Model UMA60F-7R5, Input rated 100-240VAC, 1.0A, 50-60Hz, Output rated 7.5 VDC/Max 5.5A

Model UMA60F-12, Input rated 100-240VAC, 1.4A, 50-60Hz, Output rated 12 VDC/Max 4.5A

Model UMA60F-15, Input rated 100-240VAC, 1.4A, 50-60Hz, Output rated 15 VDC/Max 3.5A

Model UMA60F-24, Input rated 100-240VAC, 1.4A, 50-60Hz, Output rated 24 VDC/Max 2.5A

Model UMA60F-36, Input rated 100-240VAC, 1.4A, 50-60Hz, Output rated 36 VDC/Max 1.7A

Model UMA60F-48, Input rated 100-240VAC, 1.4A, 50-60Hz, Output rated 48 VDC/Max 1.25A

Additional information (if necessary)



- ☐ UL Solutions (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- ☒ UL Solutions (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- ☐ UL Solutions (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- ☐ UL Solutions (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2023-12-20

Signature:

Thomas Wilson